

Pre-Clean, Patient Care Area

Richard Wolf Flexible Endoscopes



- Clean scope and flush channels using water or enzymatic detergent solution and/or enzymatic sponge.



- Rinse scope and channels with water (tap or sterile).



- Secure, transport to decontamination area:
- Place in transport bag or container for safe transport to Decontamination area.
- Keep scope moist to prevent debris from drying onto scope. Transport as soon as possible.