




neodisher® PreStop



Corrosion inhibitor for the pre-treatment of surgical instruments



Spray foam

Fields of application:

- Intended for use as corrosion inhibitor for spraying onto surgical instruments in operating rooms directly after use

Performance spectrum:

- Corrosion inhibitor: prevents pitting corrosion on stainless steel instruments which can be caused by adhesive residues of blood or physiological saline solution
- Self-acting pre-cleaning already in the OR
- Suitable for surgical instruments incl. MIS instruments
- Anodised aluminium must be tested for suitability before use

Special properties:

- Prevents drying of surgical residues by keeping the instruments wet
- Facilitates subsequent cleaning
- Inhibits the growth of micro-organisms on instruments during contact time (bacteriostatic)

Application and dosage:

neodisher PreStop is applied undiluted with the foam sprayer onto the surgical instruments in the transport container for complete wetting. Keep transport container closed. The instruments treated with neodisher PreStop can be stored for up to 72 hours. Then rinse carefully and reprocess as usual. In the case of a subsequent manual reprocessing with a detergent without disinfecting efficacy, rinsing is not necessary.

General instructions for use:

- For professional use only.
- For applying neodisher PreStop the foam sprayer which is within the scope of delivery is to be used.
- Do not mix with other products.
- Reprocessing should comply with all ordinances pursuant to the regulations on medical devices and should be performed with appropriate validated processes.
- Please observe the application recommendations given by the instrument manufacturer in accordance with the requirements of public standards such as ANSI/AAMI ST 79:2017.

Technical data:

pH-value	9.5 (undiluted, 68 °F/ 20 °C)
Density	approx. 1.04 g/cm ³ (68 °F/20 °C)

Ingredients:

All components are listed in the TSCA inventory.
N-(2-hydroxyethyl)-N-[2-[(1-oxooctyl)amino]ethyl]-β-alanine;

Concentration: <1%

1,2-benzisothiazol-3(2h)-one

Concentration: <0.05%

1,2,3-benzotriazole

Concentration: <1%


CE-mark: **CE**

neodisher PreStop complies with European guidelines for medical devices.



neodisher[®] PreStop

Storage information:

Always store at a temperature between 32 °F and 77 °F/0 °C and 25 °C. Keep away from sunlight. Keep container tightly closed. Usable for 2 years when stored as recommended. For expiry date refer to the stamp mark on the label behind the hourglass symbol .

Hazard and precautionary statements:

neodisher PreStop is not a hazardous product according to the Appendix D, OSHA Hazard Communication Standard 29 CFR 1910:1200. Dispose only when container is empty and closed. For disposal of product residues, refer to Safety Data Sheet.

MB 4086/2-8 Date of Issue: 10/2021

With the above information, which is appropriate to our current knowledge we describe our product regarding possible safety necessities, but we do not involve any quality description or promise certain properties.



Manufactured by:
Chemische Fabrik Dr. Weigert GmbH & Co. KG
Mühlenhagen 85 • Germany-20539 Hamburg

Phone: +49-40-7 89 60-0 info@drweigert.com
Fax: +49-40-7 89 60-120 www.drweigert.com

Distributed by:
Richard Wolf Medical Instruments Corp.
353 Corporate Woods Parkway
Vernon Hills, IL 60061 • U.S.A.

Phone: 847-913-1113